
Performance	<b>ENGLISH</b>	SI	Notes
Voltage Sensitivity (± 20 %)	1.0 mV/g	0.1 mV/(m/s <sup>2</sup> )	
Measurement Range	± 5000 g pk	$\pm 49050 \text{ m/s}^2 \text{ pk}$	
Frequency Range (- 5 %)	2 to 4000 Hz	2 to 4000 Hz	[2]
Resonant Frequency	≥ 55 kHz	≥ 5 kHz	
Electrical Filter Roll-off	6 dB/octave	6 dB/octave	[1]
Electrical Filter Corner Frequency	20 kHz	20 kHz	[1]
Broadband Resolution (1 to 10000 Hz)	0.03 g rms	0.29 m/s <sup>2</sup> rms	[1]
Non-Linearity	≤ 2.5 %	≤ 2.5 %	[3]
Transverse Sensitivity	≤ 5 %	≤ 5 %	
Environmental			
Overload Limit (Shock)	± 10000 g pk	$\pm$ 98100 m/s <sup>2</sup> pk	
Temperature Range (Operating)	-65 to +250 °F	-54 to +121 °C	
Temperature Response	See Graph	See Graph	
Electrical	•	•	
Excitation Voltage	18 to 30 VDC	18 to 30 VDC	
Constant Current Excitation	2 to 20 mA	2 to 20 mA	
Output Impedance	≤ 200 Ohms	≤ 200 Ohms	
Output Bias Voltage	7 to 12 VDC	7 to 12 VDC	
Discharge Time Constant	1.5 to 3.0 s	1.5 to 3.0 s	
Settling Time (within 10 % of bias)	< 10 s	< 10 s	
Spectral Noise (1 Hz)	9000 μg/√Hz	88290 (µm/s²)/√Hz	[1]
Spectral Noise (10 Hz)	2500 µg/√Hz	24525 (µm/s²)/√Hz	[1]
Spectral Noise (100 Hz)	800 µg/√Hz	7848 (µm/s²)/√Hz	[1]
Spectral Noise (1 kHz)	250 µg/√Hz	2453 (µm/s²)/√Hz	[1]
Physical		,	
Sensing Element	Ceramic	Ceramic	
Sensing Geometry	Shear	Shear	
Housing Material	Titanium	Titanium	
Sealing	Hermetic	Hermetic	
Size (Height x Length x Width)	0.40 x 0.77 x 0.40 in	10.2 x 19.6 x 10.2 mm	
Weight	0.18 oz	5.0 gm	[1]
Electrical Connector	1/4-28 4-Pin	1/4-28 4-Pin	
Electrical Connector Position	Side	Side	
Mounting Thread	5-40 Female	5-40 Female	
	Typical Sensitivity Davis	ation vs Temperature	





All specifications are at room temperature unless otherwise specified In the interest of constant product improvement, we reserve the right to change specifications without notice. ICP® is a registered trademark of PCB Piezotronics, Inc.



## **Notes**

- [1] Typical
- [2] All axes filtered to provide -5 % between 4000 and 6000 Hz
- [3] Zero-based, least-square, straight line method

## **Supplied Accessories**

NIST Traceable Calibration Certificate
Model 080A90 Mounting Stud, 10-32 to 5-40 (1)



3425 Walden Ave Depew, NY 14043 www.LarsonDavis.com